

Rabbit Anti-Human CD38 Monoclonal Antibody (Clone SP149)

CATALOG #:

M4490 0.1 ml rabbit monoclonal antibody purified by protein A/G in PBS/1% BSA buffer pH 7.6 with less than 0.1%

sodium azide.

M4492 0.5 ml rabbit monoclonal

antibody purified by protein A/G in PBS/1% BSA buffer pH 7.6 with less than 0.1%

sodium azide.


M4494 1.0 ml rabbit monoclonal

antibody purified by protein A/G in PBS/1% BSA buffer pH 7.6 with less than 0.1%

sodium azide.

M4491 7.0 ml pre-diluted rabbit

monoclonal antibody purified by protein A/G in TBS/1% BSA buffer pH 7.6 with less than 0.1% sodium azide.



Human tonsil stained with anti-CD38 antibody

Flow cytometric analysis of rabbit anti-CD38 (SP149) antibody in HeLa (green) compare to negative control of rabbit IgG (blue)

INTENDED USE: For Research Use Only. Not for use in diagnostic procedures.

CLONE: SP149

IMMUNOGEN: Synthetic peptide derived from the C-terminus of human CD38 protein.

IG ISOTYPE: Rabbit IgG
EPITOPE: Not determined

MOLECULAR WEIGHT: 46 kDa

SPECIES REACTIVITY: Human (tested). (See www.springbio.com for information on species reactivity predicted by

sequence homology.)

DESCRIPTION: CD38 (cluster of differentiation 38), also known as cyclic ADP ribose hydrolase, is a transmembrane

glycoprotein found on the surface of some immune cells including plasma cells, activated or immature T and B cells, monocytes, and natural killer cells. CD38 participates in cell adhesion, signal transduction and calcium signaling. It is expressed at high levels in the plasma cell tumor, prostate cancer, stomach cancer, and neuroblastoma. CD38 is used as one of the plasma cell

markers and its ligand is CD31 molecules.

APPLICATIONS: Immunohistochemistry (IHC) and Flow Cytometry

IHC PROCEDURE: Specimen Preparation: Formalin-fixed, paraffin-embedded tissues are suitable for use with this

primary antibody.

Deparaffinization: Deparaffinize slides using xylene or xylene alternative and graded alcohols. **Antibody Dilution:** If using the concentrate format of this product, dilute the antibody 1:100. The dilutions are estimates; actual results may differ because of variability in methods and protocols. **Antigen Retrieval:** Boil tissue section in 10mM citrate buffer, pH 6.0 for 10 min followed by cooling

at room temperature for 20 min.

Primary Antibody Incubation: Incubate for 10 minutes at room temperature.

Slide Washing: Slides must be washed in between steps. Rinse slides with PBS/0.05% Tween. **Visualization:** Detect the antibody as instructed by the instructions provided with the visualization

system.

IHC POSITIVE CONTROL: Tonsil

FLOW CYTOMETRY: Recommended starting protocol: Dilute the antibody 1:100. Incubate for 30 minutes at 4°C.

The dilution is an estimate; actual results may differ because of variability in methods and protocols.

Optimal dilution and procedure should be determined by the end user.

FLOW CYTOMETRY
POSITIVE CONTROL

POSITIVE CONTROL: HeLa Cell Line CELLULAR LOCALIZATION: Membrane

4300 Hacienda Drive Pleasanton, CA 94588 www.springbio.com

Tel: 1-925-474-8440 Toll Free: 1-800-787-6896

Fax: 1-925-474-8469

Tel: 1-925-474-8440

Toll Free: 1-800-787-6896 Fax: 1-925-474-8469

STORAGE & STABILITY:

Store at 2-8°C. Do not freeze. The user must validate any other storage conditions. When properly stored, the reagent is stable to the date indicated on the label. Do not use the reagent beyond the expiration date.

There are no definitive signs to indicate instability of this product; therefore, positive and negative controls should be tested simultaneously with unknown specimens.

If unexpected results are observed which cannot be explained by variations in laboratory procedures and a problem with the reagent is suspected, contact Technical Support at spring.tech@ventana.roche.com.

WARNINGS & PRECAUTIONS:

- Avoid contact of reagents with eyes and mucous membranes. If reagents come into contact with sensitive areas, wash with copious amounts of water.
- 2. This product is harmful if swallowed.
- 3. Consult local or state authorities with regard to recommended method of disposal.
- 4. Avoid microbial contamination of reagents.